

N.Y.S. Approved ELAP
ID: 10708

Converse Laboratories, Inc.
800 Starbuck Ave. Suite B101
Watertown, NY 13601
(315) 788-8388

U.S.P.H. Certified
36144

* Laboratory Report Form *

Sylvia Lake Association
1 Indian Head Trail
Gouverneur, NY 13642

Client ID 7607185
Attention: William deLorraine
Report Date 08/01/2018

Sample ID: 01807524 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1017 Sample Site: Hotel Beach
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.

Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	<1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	66.3 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ

Sample ID: 01807525 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1019 Sample Site: Robinsons Bay
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.

Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	<1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	159.7 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ



Sample ID: 01807526 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1030 Sample Site: Parkers Bay
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.



Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	<1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	63.7 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ

Key: mg/L - Milligrams Per Liter
ml/L - Milliliters Per Liter
100 ml - Size of Coliform Container
CFU/ml - Colony Forming Units per Milliliter
ND - None Detected
TNTC - Too Numerous to Count

All times shown in 24 hour format

E - Estimated Value


Authorized Review



Supervisor 

The information in this report is accurate to the best of our knowledge and ability.
In no event shall our liability exceed the cost of these services.
I certify that these results conform to New York State Department of Health Standards and requirements
(10 NYCRR Subpart 55 - 2).

Sample results are based on samples as they are received, unless sampled by Converse
Laboratories, Inc. This report shall not be reproduced, except in full, without written
Approval from Converse Laboratories, Inc.

N.Y.S. Approved ELAP
ID: 10708

Converse Laboratories, Inc.
800 Starbuck Ave. Suite B101
Watertown, NY 13601
(315) 788-8388

U.S.P.H. Certified
36144

* Laboratory Report Form *

Sylvia Lake Association
1 Indian Head Trail
Gouverneur, NY 13642

Client ID 7607185
Attention: William deLorraine
Report Date 08/01/2018

Sample ID: 01807521 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1005 Sample Site: Pumphouse Bay
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.

Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	<1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	193.5 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ

Sample ID: 01807522 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1010 Sample Site: Outlet Creek
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.

Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	<1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	66.4 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ


Sample ID: 01807523 Sample Type: Lake Water
Sample Date: 07/31/18 Sample Time: 1013 Sample Site: Old Beach
Received Date: 07/31/18 Received Time: 1200 Sampled By: W.D.

Analysis	Results	Method Code	Lab ID	Date	Time	Tech
E. COLI	1.0 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ
TOTAL COLIFORM	118.7 MPN/100ml	SM-21-9223B	10708	7/31/2018	1233	DKZ

Key: mg/L - Milligrams Per Liter
ml/L - Milliliters Per Liter
100 ml - Size of Coliform Container
CFU/ml - Colony Forming Units per Milliliter
ND - None Detected
TNTC - Too Numerous to Count

All times shown in 24 hour format

E - Estimated Value



Omara K. Zang
Supervisor

The information in this report is accurate to the best of our knowledge and ability.
In no event shall our liability exceed the cost of these services.
I certify that these results conform to New York State Department of Health Standards and requirements
(10 NYCRR Subpart 55 - 2).

Sample results are based on samples as they are received, unless sampled by Converse
Laboratories, Inc. This report shall not be reproduced, except in full, without written
Approval from Converse Laboratories, Inc.

[illegible]

Doc. # 357

1/19/2015

Rev. # 115

Page 1 of 1

Amt. Due:

Amt. Paid:

Initial Review:

Transcriptual Rev.:

Final Review: